

TriCorn Reporting - ISO Quality System - Calibrations Detail

<i>Equipment Reference</i>	<i>Description</i>	<u>CALIBRATIONS</u>				<u>CHECKS</u>				<i>Status</i>
		<i>Period</i>	<i>Interval</i>	<i>Last</i>	<i>Next</i>	<i>Period</i>	<i>Interval</i>	<i>Last</i>	<i>Next</i>	
ADD 101 INS	0-25mm Micrometer	Months	12	19/06/2003	19/06/2004	Months	2	05/05/2003	05/07/2003	Active
<i>Manufacturer</i> Moore & Wright										
<i>Location</i> Inspection Dept										
<i>Purchased</i> 12/03/2002 <i>First Active</i> 05/05/2002		Due to be replaced with new instrument at the end of 2003								

	<i>Date Due</i>	<i>Date Done</i>	<i>Performed By</i>	<i>Within Tolerance</i>	<i>Certificate Number</i>	<i>Adjusted</i>	<i>Actual Size</i>
Calibration	05/03/2003	05/03/2003	ADD	Yes	HJ4566	No	25
<i>Notes</i>	Ensure item is marked with ID number						
	05/04/2003	05/04/2003	ADD LTD	Yes	GH5666	No	25.001
<i>Notes</i>							
	05/05/2003	05/05/2003	ER LTD	Yes	BE2333	Yes	25.001
<i>Notes</i>	Slight mark to body.						
	22/02/2007	19/06/2003	ER LTD	Yes	EXT345	Yes	25.0001
<i>Notes</i>	Passed.						